



# Advisory Statement A17/01

## Introduction of Adult Diagnostic Reference Levels for Nuclear Medicine

### Purpose

To provide information about the introduction of adult diagnostic reference levels (DRLs) for nuclear medicine and clarify the requirements for annually comparing facility reference levels to the newly established DRLs.

### Issue

In 2017, the Australian Radiation Protection and Nuclear Safety Agency (ARPANSA) introduced adult DRLs for nuclear medicine, positron emission tomography (PET) and the computed tomography (CT) portion of single-photon emission computed tomography (SPECT)/CT and PET/CT. These DRLs were developed in collaboration with the Australian and New Zealand Society for Nuclear Medicine (ANZSNM), the Australasian Association of Nuclear Medicine Specialists (AANMS) and the Australasian College of Physical Scientists and Engineers in Medicine (ACPSEM). The [DRLs](#), which are published on the ARPANSA website, came into effect on 1 July 2017.

*Standard 3.2, Optimised Radiation Technique Charts Standard* of the DIAS requires a practice providing services which use ionising radiation to establish a program to ensure that radiation doses administered to patients are

- annually compared with diagnostic reference levels (DRLs) for diagnostic imaging procedures for which DRLs have been established in Australia; and
- if DRLs are consistently exceeded, a review is undertaken to determine whether radiation protection has been optimised.

The introduction of adult DRLs for nuclear medicine require practices providing Medicare funded nuclear medicine imaging services, including PET and SPECT services, to undertake an annual comparison of their facility reference levels against the Australian DRLs.

### Requirements

While the DRLs came into effect on 1 July 2017, practices providing nuclear medicine imaging services were not required to submit audit records relating to the newly established DRLs until **1 July 2018**.

During the 12 months to 1 July 2018, practices were required to review their dose metrics and facility reference levels and include in their practice policy a procedure for conducting annual audits against the newly established nuclear medicine DRLs. From 1 July 2018, as part of an application for accreditation (including a renewal of accreditation), accreditors require evidence of up to date practice policies and DRL related audit records for Medicare funded nuclear medicine imaging and any other procedures performed at the practice for which DRLs have been established.

Where a practice provides general nuclear medicine imaging services and PET services (where weight correction is not used) existing imaging protocols need to include the relevant adult DRLs. If the doses specified in the imaging protocol are higher than the established DRL, a justification for the higher dose must be provided.

For the CT component of multi-modality imaging, a practice should ensure that, for commonly performed examinations, the Dose Length Product (DLP) delivered to a sample of patients is recorded, and the median DLP compared to the DRL. A suggested sample size is between 10 and 20 patients. Where a practice performs very low volumes of a particular scan there is no requirement to conduct a survey for that procedure. A similar audit should be performed in cases where a practice administers a dose that is dependent on the weight of the patient. Where a weight based DRL is used (e.g. FDG PET scans), the proportion of patients who received a dose below the DRL should be calculated. In cases where there is a single DRL (e.g. Myocardial Perfusion Imaging studies) the median dose delivered at the facility should be compared to the DRL.

There are no DRLs for paediatric nuclear medicine imaging scans. Consequently, there is no requirement to conduct annual DRL audits of paediatric protocols.

Practices should develop their own program for acquiring the data, calculating their own facility reference levels and comparing them to the established DRLs. [See the ARPANSA website for resources](#) that may assist.

### **For more information**

Enquiries relating to the calculation of DRLs and the operation of the National Diagnostic Reference Level Service should be directed to ARPANSA by calling 1800 033 972 or by emailing [ndrld@arpansa.gov.au](mailto:ndrld@arpansa.gov.au).

Practices should also [refer to the DIAS User Guide for Practices Applying for Accreditation for clarifying information](#) about the evidentiary requirements for Standard 3.2, and contact their accreditor for further information and advice.

# Diagnostic Imaging Accreditation Scheme Advisory Statement A17/01

<b>Title</b>	<b>Introduction of Adult Diagnostic Reference Levels for Nuclear Medicine</b>
<b>Version</b>	Version 2
<b>Date of publication</b>	July 2017
<b>Replaces</b>	Version 1
<b>Status</b>	Active
<b>Compliance</b>	Mandatory
<b>Approved by</b>	Standard 3.2 Working Group, Diagnostic Imaging Accreditation Scheme Advisory Committee
<b>Last review date</b>	May 2020
<b>Next review date</b>	May 2022
<b>Information in this statement applies to</b>	DIAS Accreditors Providers of Medicare-funded nuclear medicine imaging services
<b>Relevant standard</b>	Standard 3.2, Optimised Radiation Technique Charts
<b>Prepared by</b>	Secretariat, Diagnostic Imaging Accreditation Scheme (DIAS) Advisory Committee
<b>Contact details</b>	Phone: 02 6289 8859 Email: <a href="mailto:dias@health.gov.au">dias@health.gov.au</a>
<b>Trim reference</b>	D20-1435905
<b>Links to other statements or advisory documents</b>	User Guide for Practices Applying for Accreditation
<b>Notes (if applicable)</b>	N/A